

SERIAL NUMBER

Every HOFFRICHTER GmbH device is supplied with a serial number for traceability purposes.

Please enter your device's serial number here. You will find the serial number on the rating plate on the bottom of the device.

Serial number:

Please always quote the serial number for all queries and complaints.

CONFORMITY




The device complies with the requirements of Directive 93/42/EEC.

CONTENTS

Scope of delivery	5
General.....	6
Information on user's manual	6
Symbols on the rating plate.....	6
Symbols used in this user's manual.....	7
Safety information	8
General safety instructions	8
Electrical safety	10
Installation requirements and transport	11
Instructions before commissioning	12
Using oxygen.....	12
Intended use.....	13
Contraindications.....	14
Side effects.....	15
Description of device.....	16
Housing, display and control elements	16
Technical description.....	18
Power supply.....	19
TREND II CPAP	19
TREND II AUTO	20
TREND II BILEVEL	21
TREND II BILEVEL ST20 and ST30	22
Using the device	23
Commissioning	23
Ending therapy.....	24
Switching off the device	24
Standby mode	24
Using oxygen.....	25
Device functions	26
Info-Menu	27
Patient-Menu	33
Clinic-Menu	41

Alarm functions of the device	54
Power failure	54
Mask alarm	54
Pressure alarm	55
Wake alarm	55
Low respiratory minute volume (only ST30)	55
Apnea alarm (only ST30 in mode S)	55
Using an SD card	56
General information	56
Inserting the SD card	56
Removing the SD card	57
What is saved on the SD card	57
Error messages	59
Important display messages	61
Changing the filter cassette, cleaning	63
Changing the filter cassette	63
Cleaning the mask	64
Cleaning of the therapy tube	64
Cleaning the device	65
Cleaning the head gear	65
Cleaning the humidifier	65
Troubleshooting	66
Maintenance	68
Preparing the device when changing patient	68
Disposal	69
Device	69
Packaging	69
Accessories	69
Accessories	70
Technical data	71
Manufacturer's declaration on electromagnetic compatibility	74
Disclaimer	78

SCOPE OF DELIVERY



GENERAL

INFORMATION ON USER'S MANUAL

Read this user's manual through carefully before using the therapy device for the first time.

Follow the safety and cleaning instructions in particular.

Keep the manual in a safe place close to the device so that you can refer to it immediately if necessary.

SYMBOLS ON THE RATING PLATE



Observe the warning and safety instructions in the user's manual.



BF application part



Protection class II (protective insulation)



CE conformity declaration



Follow the user's manual.



Manufacturer



Do not dispose of the device in the household waste. Please contact the relevant customer services department to find out how to dispose of the device properly.

SYMBOLS USED IN THIS USER'S MANUAL

Important information is denoted by symbols in this user's manual. Be sure to follow these instructions in order to avoid accidents, personal injury and material damage.

In addition, the local accident prevention regulations and general safety regulations in force in the area of use must be observed.



This symbol denotes general safety instructions. Follow these instructions to avoid accidents, personal injury or material damage.

DANGER

This symbol denotes hazardous situations that lead to serious injuries or death.

WARNING

This symbol denotes hazardous situations that may lead to serious injuries or death.

CAUTION

This symbol denotes hazardous situations that may lead to moderately severe injuries.

NOTICE

This symbol denotes situations that may lead to material damage or damage to the device.

IMPORTANT

This symbol denotes information, tips and instructions for the efficient, error-free use of the device.

SAFETY INFORMATION

GENERAL SAFETY INSTRUCTIONS



- The device must only be used for your own CPAP therapy, which is prescribed by the physician.
- Use only accessories and spare parts tested and approved by the manufacturer.
- For operation and usage of the device at home, a respiratory mask specially adapted for the individual patient is required in addition to the therapy device with power cable and therapy tubing system. Only use masks that have been prescribed for your treatment by a physician.
- Only use the mask together with a respiratory therapy device and when the device is switched on.
- Unless specified otherwise, the mask and therapy tube must only be used by one patient.
- Only use the mask after instruction by a qualified medical person and clarify in particular the intake of medicines and possible contraindications and side effects associated with the use of the mask.
- If you are using oxygen with the CPAP therapy, observe the fire protection and safety regulations for using oxygen.
- Do not wear the mask if the therapy device is switched off because this could lead to breathing in air that has already been exhaled. Observe the respiratory mask manufacturer's instructions.
- Always ensure that the small hole in the mask or in the adapter between the mask and the tube is unobstructed so that the exhaled air that is loaded with CO₂ can escape.
- Observe the mask manufacturer's instructions.



- Please see your physician immediately if dryness of the mucous membranes in the nose and throat, sinus discomfort, ear ache, runny nose, over sensitive reactions of the skin, irritability, loss of voice, orientation or memory impairment occur when using the device.
- Inform your specialist dealer immediately if the device is not working properly.
- Check the device's alarm functions regularly (see page 54).

ELECTRICAL SAFETY



- Do not use the device if the housing or the cable of the device or the power supply are damaged.
- Always remove the power plug before cleaning.
- Do not open the device housing under any circumstances. Inform your specialist dealer if the device develops a fault.
- The use of accessories or power supplies not approved by us for the device may lead to increased emission of electromagnetic radiation or reduced resistance to interference.
- Protect the device from water and dampness.
- Never operate the device in damp places or in a bathroom.
- Do not put any containers full of liquid on the device.
- Do not reach for the device under any circumstances should it fall into water. Immediately unplug the power plug.
- Do not put the device near water containers (baths).

INSTALLATION REQUIREMENTS AND TRANSPORT



- Place the device near the bed. Ideally a firm, level surface on your bedside cabinet is suitable.
- Never put the device on a cupboard shelf or under the bed.
- Do not put the device with its back to a wall. Ensure that the air inlet is not covered.
- Ensure that the device is in a horizontal position, especially if it is connected to a humidifier.
- Ensure that the air can circulate properly round the device. Air circulation, particularly under and behind the device, must not be obstructed by furniture, curtains or cushions.
- Place the device in such a way that the mains plug is easily accessible so that it can be unplugged quickly in the event of a hazard.
- Do not put the device close to a source of heat.
- Place the device and accessories in such a position that they are not exposed to direct sunlight.
- The device must never be put near other devices or equipment such as defibrillators, diathermy units, mobile phones, microwaves, remote controlled toys, etc. Electromagnetic fields that exceed 3 V/m may adversely affect the operation of the TREND II.
- When used as a mobile unit, it must also be placed on a level, firm surface that has a raised edge so that the device cannot fall.
- During mobile use of the device, only operate it when the vehicle is stationary.

INSTRUCTIONS BEFORE COMMISSIONING



- Do not switch the device on if it has previously been in a very cold environment. Wait about 1 hour until the temperature has balanced out.
- Check the filter cassette regularly. Change the filter cassette as described on page 63. Never use the device without the filter cassette.
- Clean your mask system regularly and check all accessories, particularly the therapy tube, mask and headgear. When doing this, observe the manufacturer's safety and cleaning instructions.

USING OXYGEN



- Oxygen supports combustion. Therefore, observe the fire protection regulations applicable for using oxygen.
- Ensure that there is no grease on the oxygen fittings. Do not smoke and do not handle naked flames.
- Before the device is used, appropriate training must be carried out in the home environment.
- Please be sure to observe the user's manual of the manufacturer or distributor from whom you obtain the oxygen.
- Have your distributor advise you about the use of oxygen.
- In any case, follow your physician's instructions.

INTENDED USE

The TREND II is used to treat obstructive sleep apnea in patients weighing more than 30 kg.

The device creates continuous, positive airway pressure to keep the upper respiratory tracts open.

The device is not suitable for use with patients undergoing artificial respiration.

DANGER

This therapy device is not a life-supporting system.

The therapy pressure is administered via a respiratory mask which must be fitted with an exhalation valve to ensure that the exhaled air is discharged.

NOTICE

If an AquaTREND uni humidifier is connected to the device, the water in the humidifier must not contain any additives such as medicines, salts, aromatic oils or other substances.

CONTRAINDICATIONS

WARNING

Respiratory therapy may be contraindicated for certain pre-existing conditions. Therefore, always talk to the physician treating you before starting the therapy.

Pre-existing conditions include:

- bullous lung diseases
- pneumothorax
- very low blood pressure
- pneumocephalus after open craniocerebral injury or other head injuries

In the event of inflammation of the paranasal sinuses or a middle-ear infection, it should be checked whether the therapy can continue. Please speak to your physician about this.

SIDE EFFECTS

There is the possibility of undesirable side effects occurring with respiratory therapy. Reasons for side effects occurring could be unsuitable therapy settings, not using the device properly or not following the cleaning instructions.

Normally the side effects disappear when the causes have been eliminated.

You will find suitable counter measures for some side effects in the section “Troubleshooting” on page 66.

The following side effects may occur during therapy:


- Pain in the nose, the paranasal sinuses and the ears
- Dryness and irritation in the nose and throat
- Nose bleed, runny nose, sneezing, colds
- Irritated or dry eyes
- Reddening of the skin, swelling of the skin and pressure points in the mask area
- Difficulty breathing, claustrophobia
- Stomach problems because of air accumulating in the stomach


CAUTION

Please contact your physician for an explanation of the causes should these complaints persist.

DESCRIPTION OF DEVICE

HOUSING, DISPLAY AND CONTROL ELEMENTS





TECHNICAL DESCRIPTION

For all devices of the TREND II series the following characteristics apply.

To generate the necessary pressure, the device is provided with an electronically controlled blower. In order to keep the stress for the patient as low as possible, the blower has been designed with a high reserve capacity and a quick control response.

The pressure is measured directly in the mask and constantly controlled by the device. Consequently, the breathing work for the patient is as easy as possible.

There are two separate tubings in the device. One is for respiratory air and one is for the convection air required for cooling the electronics.

The device contains a built-in microcontroller which controls all of its functions.


A quartz alarm clock further adds to the comfort when using the device. Further comfort functions include a soft start ramp, the automatic start-stop function and the adjustable display brightness. While passing through the device, the air is heated a little, so that it has a higher water absorbency. When indoor air is dry in winter, the mucous membranes in mouth and nose may become particularly dry. This is unpleasant and may even cause infections in exceptional cases. For that reason, a therapy might require air humidification.

AquaTREND uni a clip-on humidifier is available as accessory for all TREND II devices.

If necessary, just click the humidifier onto the device. Remove it by pressing its release button.

The heater levels are set in the heater menu. The selected heater level is displayed as in the display of the device while the humidifier heating is switched on.

For more information please see your AquaTREND uni - User's manual or contact your specialized dealer.



Press ON/OFF!
Heat : 2 21:45

POWER SUPPLY

Internal switched-mode power supply

For mains operation the TREND II has an internal switching power supply with a wide input range of 100 - 240 V alternating current (AC), 50 - 60 Hz. Thus it is possible to connect it to an energy supply anywhere in the world.

DC vehicle cable

For mobile use of the TREND II, e.g., in a truck or a caravan, it is possible to operate it with a 24 V direct current (DC). For this you will need the optionally available DC vehicle cable.

WARNING

Only use the optional DC vehicle cable to connect the device to a DC power supply and do not connect the device to a PC while it is being used.

TREND II CPAP

The CPAP device is a therapy device for treating obstructive sleep apnea. For this purpose the device provides a continuous positive air pressure. Apart from the conventional CPAP-mode the device offers a proportional increase in pressure in the FLEXLINE mode to the individual breath flow during inspiration and a pressure drop during the expiration. This intelligent method of pressure adjustment makes a careful therapy possible.

TREND II AUTO

Within a certain pressure range the device automatically controls the therapy pressure depending on the occurrence of respiratory events (APAP = Adaptive Positive Airway Pressure). In sleep phases with no events the therapy pressure is lowered. Starting obstructive events are treated effectively with a slight pressure increase.

For a high therapy comfort the therapy pressure should always be only as high as necessary and as low as possible.

The device can recognize the following respiratory events:

- Obstructive apnea
- Hypopnea
- Central apnea
- Snoring
- Normal respiration
- Increased airway resistance

The identification of the different respiratory events is achieved via 3 channels:

1. Measurement of any constriction of the respiratory tract
2. Detection of flow limitations by measuring the flow
3. Detection of snoring with the help of sound detection

The combined analysis of all results facilitates a very precise differentiation between obstructive and central events.

The device adjusts the therapy pressure according to the detected events as described below:

- Increase of pressure for obstructive events
- Decrease of pressure for central events and normal respiration

In addition, the pressure increase can be adjusted in 5 levels according to special requirements and to the individual patient comfort. These settings are performed by the physician with the help of the PC software TRENDset. In addition to normal APAP and CPAP mode the device offers the FLEXLINE mode, a pressure increase which is proportionate to the individual respiration flow during inspiration and a pressure decrease during expiration.

TREND II BILEVEL

The device has two different adjustable pressure levels. A higher level for inspiration and a lower level for expiration.

Trigger device

The respiration trigger detects the patients efforts to breath in and out and signals these efforts to the control device. A volume-based trigger can be set for inspiration, and a flow-based trigger for expiration. In conjunction with an optimum slope adjustment, this results in a therapy which is very comfortable for the patient and which gives the impression as if the device is following the natural respiration without any delay.

NOTICE

During the expiration phase a negative pressure (vacuum) will not be created.

CPAP-mode

The device provides continuous positive pressure for the therapy.

S-mode

Spontaneous trigger – the change from one pressure level to another is triggered by starting respiration of the patient only. When the inspiration flow ends, the pressure drops to the lower level set. In addition, a backup frequency (Backup freq) can be activated so that in the event of apnea, the patient continues to be ventilated with fixed parameters.

TREND II BILEVEL ST20 AND ST30

The device has two different adjustable pressure levels. A higher level for inspiration and a lower level for expiration.

Trigger device

The respiration trigger detects the patients efforts to breath in and out and signals these efforts to the control device. A volume-based trigger can be set for inspiration, and a flow-based trigger for expiration. In conjunction with an optimum slope adjustment, this results in a therapy which is very comfortable for the patient and which gives the impression as if the device is following the natural respiration without any delay.

The modes CPAP, S, ST and T are integrated in the device.

CPAP-mode

The device provides continuous positive pressure for the therapy.

S-mode

Spontaneous trigger – the change from one pressure level to another is triggered by starting respiration of the patient only. When the inspiration flow ends, the pressure drops to the lower level set. In addition, a backup frequency (Backup freq) can be activated so that in the event of apnea, the patient continues to be ventilated with fixed parameters.

ST-mode

Spontaneous breathing and Timed – as long as the patient is breathing spontaneously the device is operating in the S-mode. If the patient stops breathing, the device automatically switches to the T mode and ventilates the patient with the set frequency. The delay time after which the device switches into the T-mode is adjustable.

T-mode

Timed respiration with prescribed, set inspiration time. If the trigger function is deactivated the patient is ventilated according to the set parameters. If the trigger function is activated the device accepts only spontaneous trigger signals in the expiration phase.

NOTICE

During the expiration phase a negative pressure (vacuum) will not be created.

USING THE DEVICE

COMMISSIONING

⚠ CAUTION


Before commissioning the device, read the section “Safety information” (from page 8).

The device must not be operated without a filter cassette because this could damage your health.

1. After having set up the device properly, connect the device using the supplied power cord, with the socket outlet.
2. The device displays a welcoming message, the software version and the current number of treatment hours. Thereafter, the display shows the date and time.
3. Connect the therapy tube
 - a) to the air outlet, and
 - b) to the mask.
4. Also connect the measuring tube to the measuring tube connection.
5. Put on the mask. If you have selected the automatic mode, the device will start on your first breath. In the manual mode, press the on/off key to start the device.
6. The device first of all runs at maximum pressure for the time selected by you for the mask test. You should now properly seat the mask in order to eliminate any potential leaks.
7. Place the tube such that it does not exert any pulling action on the mask while you are lying down
8. Breathe deeply and calmly. If you have activated the soft start function, the device initially reduces the pressure after completion of the mask test and will then slowly increase the pressure automatically to the prescribed value, allowing you to fall asleep at a lower pressure.

IMPORTANT

In the section “Device functions” (from page 26) read how you can adapt the times for soft start and the mask test to suit your personal requirements.



ENDING THERAPY

To end the therapy, press the on/off key.

If you are using an SD card, the therapy data is saved on the SD card after the therapy is ended. During this time, the message “Do not remove the SD card” is shown on the display.

A black rectangular LCD display with white pixelated text that reads "Do not remove the SD card".

Do not remove
the SD card

NOTICE

On no account remove the SD card during saving. This could lead to loss of data or damage to the SD card.

SWITCHING OFF THE DEVICE

Ensure that the device is in standby mode. Then unplug the power plug.

STANDBY MODE

If the turbine is switched off, the device switches to standby mode (discernable on the display with date and time).

The TREND II can be kept in standby mode permanently. This does not harm it.

A black rectangular LCD display with white pixelated text that reads "Press ON/OFF!" on the first line and "25.May.11 12:20" on the second line.

Press ON/OFF!
25.May.11 12:20

USING OXYGEN

⚠ WARNING

Before using oxygen, it is essential to read the safety information on page 12.



When feeding oxygen directly into the mask, please use a kink-resistant tube made of a medically approved material.

Oxygen can also be fed in via an adapter fitted onto the air outlet.

Proceed as follows when using oxygen:

⚠ WARNING

If the device is in standby mode or switched off, the oxygen supply must always be switched off.



1. Before starting the treatment, check that the tube connections are fitted correctly.
2. First of all, switch the device on and then the oxygen supply.
3. Check whether the “Auto off” automatic mode is active. If not, program it as described on page 38.
4. Switch off the oxygen supply before switching the device off.

DEVICE FUNCTIONS

There are three keys for programing the TREND II functions.

- ▲ = Programing key
- ▼ = Programing key
- ◀ = Enter key

The display has two lines. Using the programing keys ▲ and ▼, you can select the line to be displayed. The triangular symbol ▶ preceding the line indicates that this line has been selected.

IMPORTANT

If you do not press any key for 30 seconds while programing, the programing mode is automatically exited for safety reasons.

Unless the changed value is confirmed with the enter key when the programing mode is exited, the original value remains as it is.

If you select the line **Exit Menu** and then press the enter key ▶, programing is ended.

The device has three different menus:

1. **Info-Menu**
In this menu the set parameters can be displayed.
2. **Patient-Menu**
In this menu the patient can change settings.
3. **Clinic-Menu**
In this menu, only a physician or service technicians can change settings.

Use the programing key ▲ for scrolling within the menus. The sequence of the menu structure is shown in the table from page 27.



```
▶Press.....7.5hPa
Exit Menu
```

Example: Pressure

INFO-MENU

By pressing the info key **i** you get to the Info-Menu. By pressing the programing keys **▲** and **▼** you can see the next parameters. For the parameters “Use”, “AI” or “AHI” and “Leak”, pressing the enter key **◀** enables you to display average values across different periods of time. To exit the Info-Menu select the line **Exit Menu** and then press the enter key **◀**. If you do not press any keys for 30 seconds the device will automatically switch back to the operating mode.

The following table shows which parameters are displayed in the Info menu, depending on the set mode or device type.

Parameter	Mode					
	CPAP	FLEX	APAP	BILEVEL S	BILEVEL ST	BILEVEL T
Use[D] ▶◀						
Use[W] ▶◀						
Use[M] ▶◀	•	•	•	•	•	•
Use[6M] ▶◀						
Use[Y] ▶◀						
AI[D] ¹ ▶◀						
AI[W] ¹ ▶◀						
AI[M] ¹ ▶◀	•	•	•	•	•	•
AI[6M] ¹ ▶◀						
AI[Y] ¹ ▶◀						
Leak[D] ▶◀						
Leak[W] ▶◀						
Leak[M] ▶◀	•	•	•	•	•	•
Leak[6M] ▶◀						
Leak[Y] ▶◀						
Therapy	•	•	•	•	•	•
Filter	•	•	•	•	•	•
Ramp	•	•	•	•	•	•
P-Ramp	•	•	•	•	•	•
Mode	•	•	•	•	•	•
Press.	•	•	-	-	-	-
I-Pres.	-	-	-	•	•	•
E-Pres.	-	-	-	•	•	•

Parameter	Mode					
	CPAP	FLEX	APAP	BILEVEL S	BILEVEL ST	BILEVEL T
FLEX-Level	-	●	-	-	-	-
P-Start	-	-	●	-	-	-
P-Min	-	-	●	-	-	-
P-Max	-	-	●	-	-	-
I-Slope	-	-	-	●	●	●
E-Slope	-	-	-	●	●	●
Trigger	-	-	-	-	-	●
I-Trigger	-	-	-	●	●	● ⁴
E-Trigger	-	-	-	●	●	-
Frequency	-	-	-	-	●	●
Backup freq	-	-	-	●	-	-
I:E	-	-	-	-	●	●
Delay Time	-	-	-	-	●	-
VT min	-	-	-	● ³	● ³	● ³
P-addit.	-	-	-	● ⁵	● ⁵	● ⁵
TI min	-	-	-	● ⁶	● ⁶	-
TI max	-	-	-	● ⁶	● ⁶	-
Low MV ²	-	-	-	● ³	● ³	● ³
Apnea ²	-	-	-	● ^{3, 7}	-	-

◀ Next parameter is displayed by pressing the enter key

1 in TREND II AUTO “AHI” (Apnea-Hypopnea-Index) instead of “AI” is displayed

2 only available when activated by physician

3 only TREND II BILEVEL ST30

4 only if trigger is “ON”

5 cannot be set if VT min is “OFF”

6 only TREND II BILEVEL ST20 and ST30

7 only if Backup frequency is “OFF”

MEANING OF THE PARAMETERS IN THE INFO MENU

►Use[D].....8.0h
Exit Menu

Use[D]

Shows the average operating time of the last therapy day¹

►Use[W].....7.5h
Exit Menu

Use[W]

Shows the average operating time, per therapy day¹, in the last week

►Use[M].....8.0h
Exit Menu

Use[M]

Shows the average operating time, per therapy day¹, in the last month

►Use[6M].....8.5h
Exit Menu

Use[6M]

Shows the total operating time, per therapy day¹, in the last 6 months

►Use[Y].....8.5h
Exit Menu

Use[Y]

Shows the average operating time, per therapy day¹, in the last year

►AHI[D].....9.2
Exit Menu

AI[D]²

Shows the apnea index of the last night

►AI[W].....5.9
Exit Menu

AI[W]²

Shows the apnea index of the last week

►AI[M].....6.5
Exit Menu

AI[M]²

Shows the apnea index of the last month

►AI[6M].....6.8
Exit Menu

AI[6M]²

Shows the apnea index of the last 6 months

►AI[Y].....6.4
Exit Menu

AI[Y]²

Shows the apnea index of the last year

¹ from 12.00 to 12.00 o'clock

² in TREND II AUTO "AHI" (Apnea-Hypopnea-Index) instead of "AI" is displayed

Leak[D]

Shows the percentage of the operating time in which increased leakage was detected, over the period of the last therapy day¹

```
▶Leak[D].....5%  
Exit Menu
```

Leak[W]

Shows the percentage of the operating time in which increased leakage was detected, over the period of the last week

```
▶Leak[W].....2%  
Exit Menu
```

Leak[M]

Shows the percentage of the operating time in which increased leakage was detected, over the period of the last month

```
▶Leak[M].....4%  
Exit Menu
```

Leak[6M]

Shows the percentage of the operating time in which increased leakage was detected, over the period of the last 6 months

```
▶Leak[6M].....3%  
Exit Menu
```

Leak[Y]

Shows the percentage of the operating time in which increased leakage was detected, over the period of the last year

```
▶Leak[Y].....3%  
Exit Menu
```

Therapy

Shows the total therapy time of the device

```
▶Therapy.....123h  
Exit Menu
```

Filter

Shows the time since the last filter change

```
▶Filter.....123h  
Exit Menu
```

Ramp

Shows the remaining time of the soft start ramp

```
▶Ramp.....50min  
Exit Menu
```

P-Ramp

Shows the ramp start pressure

```
▶P-Ramp.....4.5hPa  
Exit Menu
```

Mode

Shows the selected mode

```
▶Mode.....Bilev ST  
Exit Menu
```

¹ from 12.00 to 12.00 o'clock

►Press.....7.5hPa
Exit Menu

Press. (only in modes CPAP and FLEX)
Shows the selected therapy pressure

►FLEX-Level.....3
Exit Menu

FLEX-Level (only in mode FLEX)
Shows the selected FLEX-Level

►P-Start....7.5hPa
Exit Menu

P-Start (only in mode APAP)
Shows the selected start pressure

►P-Min.....5.0hPa
Exit Menu

P-Min (only in mode APAP)
Shows the selected minimum pressure

►P-Max.....10.0hPa
Exit Menu

P-Max (only in mode APAP)
Shows the selected maximum pressure

►I-Pres...10.0hPa
Exit Menu

I-Pres. (only in modes S, T and ST)
Shows the selected inspiration pressure

►E-Pres....6.0hPa
Exit Menu

E-Pres. (only in modes S, T and ST)
Shows the selected expiration pressure

►I-Slope.....3s
Exit Menu

I-Slope (only in modes S, T and ST)
Shows the selected inspiration slope

►E-Slope.....3s
Exit Menu

E-Slope (only in modes S, T and ST)
Shows the selected expiration slope

►Trigger.....OFF
Exit Menu

Trigger (only in mode T)
Shows if the inspiration trigger is activated or not

►I-Trigger.....2
Exit Menu

I-Trigger (only in modes S, T and ST)
Shows the selected inspiration trigger threshold

►E-Trigger.....2
Exit Menu

E-Trigger (only in modes S and ST)
Shows the selected expiration trigger threshold

Frequency (only in modes T and ST)

Shows the selected number of breaths per minute

►Frequency..12bpm
Exit Menu

Delay Time (only in mode ST)

Shows the delay for switching from S-mode to T-mode

►Delay Time.....3s
Exit Menu

Backup freq (only in mode S in TREND II BILEVEL, BILEVEL ST20 and ST30)

Shows whether the backup frequency is switched on or not

►Backup Freq....ON
Exit Menu

I:E (only in modes T and ST)

Shows the inspiration duration and the ratio of inspiration to expiration

►I:E....1:2.1*1.6s
Exit Menu

VT min (only ST30 in mode ST)

Shows the target volume per breath

►VT min.....200ml
Exit Menu

P-addit. (only ST30 in mode ST)

Shows the pressure that is added in order to reach the target volume

►P-addit...3.0hPa
Exit Menu

TI min (only ST30 in modes S and ST)

Shows the minimum inspiration time

►TI min.....1.2s
Exit Menu

TI max (only ST30 in modes S and ST)

Shows the maximum inspiration time

►TI max.....3.0s
Exit Menu

Low MV (only ST30 in modes S, T and ST)

Shows whether the alarm for too low respiratory minute volume “Low MV” is active or deactivated

►* Low MV.....OFF
Exit Menu

Apnea (only ST30 in mode S, if backup frequency “OFF”)

Shows whether the “Apnea” alarm is active or deactivated

►* Apnea.....OFF
Exit Menu

PATIENT-MENU

Press and hold the enter key ◀ (at least 1 second) to move to the programming mode. The Patient-Menu will now appear on the display.

The following table shows which parameters can be set in the Patient-menu, depending on the device type.

Device Parameter	CPAP	AUTO	BILEVEL	BILEVEL ST20	BILEVEL ST30
Wake Time	•	•	•	•	•
Delete Filter	•	•	•	•	•
Mask	-	•	-	-	-
Ramp	•	•	•	•	•
Mask Test	•	•	•	•	•
Auto	•	•	•	•	•
Display VT	•	•	•	•	•
Brightness	•	•	•	•	•
Date	•	•	•	•	•
Time	•	•	•	•	•

SETTING PARAMETERS AND THEIR MEANING

Wake Time

The wake time can be set or changed here.

To set the wake time, first go to the Patient-menu (hold down the enter key ◀ for approx. 1 second). Pressing the enter key ◀ causes the wake hour to flash. It can now be changed using the programing keys ▲ and ▼. After pressing the enter key ◀ again the minute of the wake time will start flashing. Now you can change it using the programing keys ▲ and ▼. Pressing the enter key ◀ again ends the input and you get back to the menu. Changing the wake time will automatically activate the alarm clock.

After leaving the menu the alarm clock can be activated and deactivated with the help of the programing keys ▲ and ▼. Press programing key ▲ to activate the alarm function. The display briefly shows the wake up time allowing you to verify it.

After that a bell symbol in front of the time display shows that the wake alarm function is activated

Press the programing key ▼ to deactivate the wake alarm function. The bell symbol disappears.

▶Wake Time..07:00
Exit Menu

Press ON/OFF!
Alarm ON 07:00

Press ON/OFF!
25.May.11 *12:20

Press ON/OFF!
Alarm Clock OFF

Delete Filter

In this menu item the filter counter can be reset. This function should be used at every filter change so that always a current value is displayed.

To reset the filter counter, go to the Patient-Menu (press the enter key ◀ key for approx. 1 second), then select **Delete Filter** with the programing keys ▲ and ▼. Press the enter key ◀. The system will now ask you whether you are sure that you want to perform a reset.

Press the enter key ◀ to perform the reset. Now the filter counter is reset to zero.

If you do not want to reset the filter counter, leave the menu using the programing keys ▲ and ▼ or wait until the system leaves the menu item automatically.

▶Delete Filter
Exit Menu



▶Mask.....Nose
Exit Menu

Mask

(only in mode APAP in TREND II AUTO)

In this menu item the type of mask can be chosen. This is important for the correct function of the therapy device. You can choose between Nose and Nose/Mouth:

- Choose **Nose** for all nasal masks and pillow masks.
- Choose **Nose/Mouth** for all nasal/mouth masks.

To select the type of mask, first go to the Patient-menu (hold down the enter key ◀ for approx. 1 second). Then select **Mask** using the programing keys ▲ and ▼ and press the enter key ◀ to change the setting.

Ramp

```
► Ramp.....50min
Exit Menu
```


Whenever the blower has been started, the microcontroller of the device checks whether a soft start ramp has been programmed. While the programmed time elapses, the soft start function increases the pressure from an adjustable initial ramp pressure (P-Ramp) up to the pressure prescribed, thus allowing you to fall asleep more easily.

In particular if you are not yet used to the respiration therapy the soft start can help you get accustomed to the device and fall asleep easily. The device allows delay times ranging from zero (the prescribed therapy pressure is applied at once) to 60 minutes (the prescribed therapy pressure is built up within 60 minutes).


The initial ramp pressure (P-Ramp) can be set by health professionals within a range from 4 hPa to the pressure prescribed.

To set the ramp, first go to the Patient-Menu (press the enter key ◀ for approx. 1 second), then select **Ramp** with the programming keys ▲ and ▼. Press the enter key ◀. Now select the desired time using the programming keys ▲ and ▼. Press enter key ◀ to confirm.

CPAP, FLEX and APAP



BILEVEL S, ST and T



►Mask Test.....30s
Exit Menu

Mask Test


This parameter determines the time when the device performs a mask test.

To ensure that the mask is tight while the pressure slowly increases in the soft start ramp even with higher pressures, you can program a mask testing time from 5 to 90 seconds in 5-second intervals. This mask test is carried out before the ramp is started. It checks whether the mask is tight even under the maximum therapy pressure.


If the FLEXLINE mode is switched on, the pressure is raised by 1 hPa at FLEX-Level 1, 2 hPa at FLEX-Level 2 and 3 hPa at FLEX-Level 3 (TREND II CPAP and TREND II AUTO). If the APAP mode is switched on, the pressure is raised to the max. pressure (TREND II AUTO) and in the devices TREND II BILEVEL and TREND II BILEVEL ST20 and ST30 the pressure in the modes S, ST and T is raised to the I-Pressure. If additional pressure is set in TREND II BILEVEL ST30, the mask test is carried out with I-Pressure + P-additional.

To set the mask test, first go to the Patient-Menu (press the enter key ◀ for approx. 1 second), then select **Mask Test** with the programming keys ▲ and ▼. Press the enter key ◀. Now select the desired time using the programming keys ▲ and ▼. Press ◀ to confirm.

CPAP, FLEX and APAP



BILEVEL S, ST and T



Auto

The automatic mode provides three settings (OFF, Start/Stop, Start):

1. Auto OFF (with mask alarm)

If you set “Auto OFF” the device must be turned on in the evening and turned off in the morning with the on/off key. If the mask slips off your face or a leak occurs while you are asleep, a visual and audible warning is emitted (mask alarm).

A black rectangular display with white text showing "Mask Alarm.....10s" on the first line and "Check Mask!" on the second line.

Mask Alarm.....10s
Check Mask!

2. Auto Start/Stop

If you set “Auto Start/Stop”, you must simply put the mask on your face. Once you start breathing, the turbine of the device is turned on. If you take the mask off, the turbine is automatically turned off after a delay time of 5 seconds. In the automatic mode, the turbine is turned off even if the mask or the tube slips off unnoticed during the night.

A black rectangular display with white text showing "Put on Mask!" on the first line and "25.May.11 22:15" on the second line.

Put on Mask!
25.May.11 22:15

3. Auto Start (with mask alarm)

The third option that can be selected is the “Auto Start” mode. Once you start breathing, the turbine of the device is turned on. If the mask slips off your face or a leak occurs while you are asleep, a visual and audible warning is emitted (mask alarm). However, the turbine is not switched off. It can only be switched off by pressing the on/off key.

A black rectangular display with white text showing "▶Auto.....Start" on the first line and "Exit Menu" on the second line.

▶Auto.....Start
Exit Menu

IMPORTANT

The automatic function can be deactivated in the TREND II BILEVEL ST20 and ST30 with the PC software from the medical staff. The device can be switched on or off in every mode by pressing the on/off key.

To set the automatic function, first go to the Patient-menu (hold down the enter key ◀ for approx. 1 second). Then select **Auto** using the programming keys ▲ and ▼. Press the enter key ◀ to change the setting.

```
►Display VT.....ON  
Exit Menu
```

Display VT

Here, you can set whether the tidal volume and the respiratory status (S=spontaneous, T=timed [only in ST mode, and in S mode when Backup frequency is "ON"]) are displayed during therapy.

To activate or deactivate the display of the tidal volume, first go to the Clinic-menu. Then select **Display VT** using the programming keys ▲ and ▼ and press the enter key ◀ to set the desired setting ("ON" or "OFF").

IMPORTANT

The display of the tidal volume during the therapy replaces the display of the date. The alarm clock is still active.

```
►Brightness..100%  
Exit Menu
```

Brightness

In this menu item you can set the brightness of the display.

To set the brightness of the display, first go to the Patient-menu (hold down the enter key ◀ for approx. 1 second). Then select **Brightness** using the programming keys ▲ and ▼. After pressing the enter key ◀ the currently set brightness will flash and can then be adjusted using the programming keys ▲ and ▼ in intervals of 10 %. Press enter key ◀ for accepting the selected brightness. If no key is pressed for 30 seconds the device will reset the brightness to the old value.

After pressing any key the brightness of the display will automatically switch to 100 % for better visibility.

Date

Use this menu item to set the date.

To set the date, first go to the Patient-menu (hold down the enter key ◀ for approx. 1 second). Then select **Date** using the programing keys ▲ and ▼. After using the enter key ◀ the year will flash and can be set using the programing keys ▲ and ▼. After using the enter key ◀ again the month will flash and can be set using the programing keys ▲ and ▼. After using the enter key ◀ again the day will flash and can be set using the programing keys ▲ and ▼. Press the enter key ◀ again to accept the set date.

A black rectangular LCD display with white pixelated text. The top line shows '▶Date...25.May.11' and the bottom line shows 'Exit Menu'.

Time

Use this menu item to set the time.

To set the time, first go to the Patient-menu (hold down the enter key ◀ for approx. 1 second). Then select **Time** using the programing keys ▲ and ▼. After pressing the enter key ◀ the hour will flash and can be set using the programing keys ▲ and ▼. After pressing the enter key ◀ again the minutes will flash and can be set using the programing keys ▲ and ▼. Press the enter key ◀ again to accept the set time.

With its built-in battery, the internal clock has a reserve power of 8 years without connection to the external power supply.

A black rectangular LCD display with white pixelated text. The top line shows '▶Time.....13:00' and the bottom line shows 'Exit Menu'.

CLINIC-MENU

In addition to the parameters in the Clinic-Menu, in the Clinic-Menu parameters relevant for the therapy can be set. This menu is used only by health and service professionals. Some parameters are protected by a PIN code in order to prevent incorrect parameters being entered.

CAUTION

Please do not try to find out the PIN. If you are in doubt as to whether the set parameters are correct, please consult your physician.

The following table shows which parameters can be set in the Clinic-menu, depending on the device type.

Parameter	Mode					
	CPAP	FLEX	APAP	BILEVEL S	BILEVEL ST	BILEVEL T
Mode ¹	•	•	•	•	•	•
Press. ¹	•	•	-	-	-	-
FLEX-Level ¹	-	•	-	-	-	-
P-Start ¹ ◀	-	-	•	-	-	-
P-Min ¹ ◀	-	-	•	-	-	-
P-Max ¹	-	-	•	-	-	-
Mask	-	-	•	-	-	-
I-Pres. ¹ ◀	-	-	-	•	•	•
E-Pres. ¹ ◀	-	-	-	•	•	•
I-Slope ¹ ◀	-	-	-	•	•	•
E-Slope ¹	-	-	-	-	-	-
Trigger ¹	-	-	-	-	-	•
I-Trigger ¹ ◀	-	-	-	•	•	• ²
E-Trigger ¹	-	-	-	-	-	-
Frequency ¹ ◀	-	-	-	-	•	•
I:E ¹ ◀	-	-	-	-	•	•
Delay Time ¹	-	-	-	-	-	-
Backup freq ¹	-	-	-	•	-	-
VT min ¹ ◀	-	-	-	• ⁵	• ⁵	• ⁵
P-addit. ¹	-	-	-	-	-	-
TI min ¹ ◀	-	-	-	• ³	• ³	-
TI max ¹	-	-	-	-	-	-

Parameter	Mode					
	CPAP	FLEX	APAP	BILEVEL S	BILEVEL ST	BILEVEL T
Ramp	•	•	•	•	•	•
P-Ramp	•	•	•	•	•	•
Mask Test	•	•	•	•	•	•
Auto	•	•	•	•	•	•
Display VT	•	•	•	•	•	•
Low MV ¹ ◀	-	-	-	• ⁵	• ⁵	• ⁵
Apnea ^{1, 4}	-	-	-	• ⁵	• ⁵	• ⁵
Brightness	•	•	•	•	•	•
Language	•	•	•	•	•	•
P-Unit	•	•	•	•	•	•
Turbine ◀	•	•	•	•	•	•
Filter ◀	•	•	•	•	•	•
Therapy ◀	•	•	•	•	•	•
Standby	•	•	•	•	•	•
Delete Filter	•	•	•	•	•	•
Date	•	•	•	•	•	•
Time	•	•	•	•	•	•
Wake Time	•	•	•	•	•	•

◀ Next parameter is displayed by pressing the enter key

1 To prevent incorrect parameters being entered, it is necessary to enter a PIN.

This PIN is known to your physician and specialist dealer

2 E-Trigger not in T mode, I-Trigger only if trigger is "ON"

3 only TREND II BILEVEL ST20 and ST30

4 Apnea only in mode S, if Backup frequency is "OFF"

5 only TREND II BILEVEL ST30

►Mode.....Bilev ST
Exit Menu

►PIN Code? 0000
Exit Menu

THERAPY-RELEVANT PARAMETERS AND THEIR MEANING

Mode

The devices can be operated in different modes. The following table shows which modes are available for your device.

To set the mode, first go to the Clinic-menu. Then select **Mode** using the programing keys ▲ and ▼. Press the enter key ◀. Now the device will prompt you to enter the PIN. The first digit is flashing. Enter the first digit of the PIN with the help of the programing keys ▲ and ▼. After pressing the enter key ◀ the second digit will flash. Enter the second digit of the PIN with the help of the programing keys ▲ and ▼, and proceed as described for the first digit. Enter the third digit accordingly. After entering the last digit of the PIN you can switch between the modes by pressing the enter key ◀. Leave the menu once the desired mode is selected. The device will then save the new mode.

Mode Device type	CPAP	FLEX	APAP	BILEVEL S	BILEVEL ST	BILEVEL T
TREND II CPAP	•	•	-	-	-	-
TREND II AUTO	•	•	•	-	-	-
TREND II BILEVEL	•	-	-	•	-	-
TREND II BILEVEL ST20	•	-	-	•	•	•
TREND II BILEVEL ST30	•	-	-	•	•	•

►Press.....7.5hPa
Exit Menu

Press (only in modes CPAP and FLEX)

To set the pressure, first go to the Clinic-menu. Then select **Pressure** using the programing keys ▲ and ▼. Press the enter key ◀. Now the device will prompt you to enter the PIN. The first digit is flashing. Enter the first digit of the PIN with the help of the programing keys ▲ and ▼. After pressing the enter key ◀ the second digit will flash. Enter the second digit of the PIN with the help of the programing keys ▲ and ▼ keys, and proceed as described for the first digit. Enter the third digit accordingly. After entering the last digit of the PIN you can set the pressure using the programing keys ▲ and ▼. Then confirm the selected pressure by pressing the enter keys ◀.

FLEX-Level (only in mode FLEX)

The value of pressure increase and pressure decrease depends on the value of the breathing flow, the set FLEX-Level and the set therapy pressure. The three FLEX-Levels (1-3) are for individual adjustment to the patients requirements:

- FLEX-Level 1 low pressure support
($\Delta P_{\max} \leq \pm 1$ hPa)
- FLEX-Level 2 medium pressure support
($\Delta P_{\max} \leq \pm 2$ hPa)
- FLEX-Level 3 high pressure support
($\Delta P_{\max} \leq \pm 3$ hPa)

To set the FLEX-Level, first go to the Clinic-menu. Then select **FLEX-Level** using the programing keys ▲ and ▼. Press the enter key ◀. Now the device will prompt you to enter the PIN. The first digit is flashing. Enter the first digit of the PIN with the help of the programing keys ▲ and ▼. After pressing the enter key ◀ the second digit will flash. Enter the second digit of the PIN with the help of the programing keys ▲ and ▼, and proceed as described for the first digit. Enter the third digit accordingly. After entering the last digit of the PIN you can set the FLEX-Level using the programing keys ▲ and ▼. Then confirm the selected level by pressing the enter key ◀.

```
►FLEX-Level.....3
Exit Menu
```

P-Start (only in mode APAP)

To set the P-Start, first go to the Clinic-menu. Then select **P-Start** using the programing keys ▲ and ▼. Press the enter key ◀. Now the device will prompt you to enter the PIN. By using the programing keys ▲ and ▼ you can now set the start pressure. Confirm with the enter key ◀. Next the parameter P-Min is set.


```
►P-Start....7.5hPa
Exit Menu
```

P-Min (only in mode APAP)

First set the P-Start. Then set the minimum pressure using the programing keys ▲ and ▼. Confirm with the enter key ◀. Next the parameter P-Max is set.

```
►P-Min.....5.0hPa
Exit Menu
```

►P-Max.....10.0hPa
Exit Menu



►I-Pres...10.0hPa
Exit Menu

P-Max (only in mode APAP)

First set the P-Start and P-Min. Then set the maximum pressure using the programing keys ▲ and ▼. Confirm with the enter key ◀.

I-Pres. (only in modes S, T and ST)

Pressure during inspiration

To set the I-Pressure, first go to the Clinic-menu. Then select **I-Press**, using the programing keys ▲ and ▼. Press the enter key ◀. Now the device will prompt you to enter the PIN. The first digit is flashing. Enter the first digit of the PIN with the help of the programing keys ▲ and ▼. After pressing the enter key ◀ the second digit will flash. Enter the second digit of the PIN with the help of the programing keys ▲ and ▼, and proceed as described for the first digit. Enter the third digit accordingly. After entering the last digit of the PIN you can set the pressure using the programing keys ▲ and ▼. Then confirm the selected pressure by pressing the enter key ◀. Next the parameter E-Pressure is set.

►E-Pres.....6.0hPa
Exit Menu

E-Pres. (only in modes S, T and ST)

Pressure during expiration

First set the I-Pressure. Then set the desired pressure using the programing keys ▲ and ▼. Press enter key ◀ to confirm. Next the parameter I-Slope is set.

►I-Slope.....3s
Exit Menu

I-Slope (only in modes S, T and ST)


Time of pressure increase from E-Pressure to I-Pressure (pressure increase rate)

First set the I-Pressure and E-Pressure. Then set the desired time using the programing keys ▲ and ▼. Press enter key ◀ to confirm. Next the parameter E-Slope is set.

E-Slope (only in modes S, T and ST)

Time of pressure decrease from I-Pressure to E-Pressure (pressure decrease rate)

First set the I-Pressure, E-Pressure and the I-Slope. Then set the desired time using the programing keys ▲ and ▼. Press enter key ◀ to confirm.



```
▶E-Slope.....3s
Exit Menu
```

Trigger (only in mode T)

To set the Trigger, first go to the Clinic-menu. Then select **Trigger** using the programing keys ▲ and ▼. Press the enter key ◀. Now the device will prompt you to enter the PIN. Pressing the enter key ◀ again changes the setting from “OFF” to “ON” and vice versa.

```
▶Trigger.....OFF
Exit Menu
```

I-Trigger (only in modes S, T and ST)

The I-Trigger depends on the volume and determines the sensitivity for detecting inspiration. It can be set to three levels.

1	10 ml
2	15 ml
3	25 ml

```
▶I-Trigger.....2
Exit Menu
```

To set the I-Trigger, first go to the Clinic-menu. Then select **I-Trigger** using the programing keys ▲ and ▼. Press the enter key ◀. Now the device will prompt you to enter the PIN. Now select the desired level by pressing the keys ▲ and ▼. Press enter key ◀ to confirm. Next the parameter E-Trigger is set.

►E-Trigger.....2
Exit Menu

E-Trigger (only in modes S and ST)

The E-Trigger depends on the respiratory flow and determines the sensitivity for detecting expiration. It can be set to three levels.

1	60 % of max. flow value
2	70 %x of max. flow value
3	80 % of max. flow value

First set the I-Trigger. Then set the desired level using the programming keys ▲ and ▼. Press enter key ◀ to confirm.

►Frequency..12bpm
Exit Menu

Frequency (only in modes T and ST)

The frequency determines the breathes per minute (only in mode T or in mode ST with apnea).

To set the frequency, first go to the Clinic-menu. Then select **Frequency** using the programming keys ▲ and ▼. Now the device will prompt you to enter the PIN. After entering the PIN set the desired frequency using the programming keys ▲ and ▼. You can set values from 5 bpm to 30 bpm in 1 bpm-intervals. Press enter key ◀ to confirm. Next the parameter I:E is set.

►I:E....1:2.1+1.6s
Exit Menu

I:E (only in modes T and ST)

The ratio I:E determines the ratio of inspiration to expiration.

First set the frequency. Then set the desired inspiration time using the programming keys ▲ and ▼. Values between 0.4 s and 8.0 s can be set in 0.1 s steps. To confirm, press the enter key ◀. The ratio I:E depends on the set inspiration time. The delay time parameter is set next.

►Delay Time.....3s
Exit Menu

Delay Time (only in mode ST)

The delay time is the time after which the device switches from S-mode into T-mode in the event of breath failure.

First set the frequency and the inspiration time. Then set the delay time using the programming keys ▲ and ▼. Press enter key ◀ to confirm. You can set values from 0 sec to 20 sec in 1 sec-intervals.

Language

The device can emit display messages in German (DEU), English (ENG), Greek (ELL), Spanish (SPA), French (FRA), Italian (ITA), Turkish (TUR), Polish (PLK) and Portuguese (POR).

To set the language, first go to the Clinic-menu. Then select **Language** using the programming keys ▲ and ▼. Press the enter key ◀. Each time you press the key another language will appear. Leave the menu once the desired language is selected. The device will then save the new language.

```
►Language.....ENG
Exit Menu
```

P-Unit

The device can show the pressure in the pressure units hectopascal (hPa), millibar (mbar), and centimeters of water (cm = cmH₂O).

To set the unit of pressure, first go to the Clinic-menu. Then select **P-Unit** using the programming keys ▲ and ▼. Press the enter key ◀. Each time you press the key another pressure unit will appear. Leave the menu once the desired pressure unit is selected. The device will then save the new unit.

```
►P-Unit.....mbar
Exit Menu
```

⚠ CAUTION

When changing the pressure unit from hPa or mbar to cmH₂O, the pressure will slightly change.

Conversion: 1 hPa \triangleq 1 mbar \triangleq 1.02 cmH₂O

Example:

At 8 cmH₂O the pressure is 7.8456 hPa or mbar.


```
▶Turbine.....123h
Exit Menu
```

```
▶Filter.....123h
Exit Menu
```

```
▶Therapy.....123h
Exit Menu
```

```
▶Standby.....123h
Exit Menu
```

```
▶Backup freq....ON
Exit Menu
```

Turbine

This menu item shows the total operating time of the turbine, the filter, and the therapy time.

Turbine

The counter indicates the total turbine running time.

Filter

The counter indicates the operating hours of the filter since the last reset.

Therapy

The counter indicates the operating time with the therapy pressure applied.

Standby

This counter indicates the time the device was connected to the power supply.

To display the counter, first go to the Clinic-menu. Then select **Turbine** using the programing keys ▲ and ▼. Press the enter key ◀. Each time you press the enter key ◀, you see the next counter.

Backup freq (only in mode S in TREND II BILEVEL, BILEVEL ST20 and ST30)

In the event of apnea lasting longer than 15 s, the patient will continue to be ventilated with a frequency of 10 bpm and an I:E ratio of 1:2.

To activate or deactivate the Backup frequency, first go to the Clinic-menu. Then select **Backup freq** using the programing keys ▲ and ▼. The PIN code is now requested. Once you have entered the PIN, you can change the setting (“ON” or “OFF”) by pressing the enter key ◀.

VT min (only in modes S, T and ST in TREND II BILEVEL ST30)

The purpose of the target volume is to safeguard the respiratory volume per breath during pressure-controlled ventilation.

To set the target volume, first go to the Clinic-menu. Then select **VT min** using the programing keys ▲ and ▼. You will now be requested to enter the PIN. The first digit will flash. Use the programing keys ▲ and ▼ to enter the first digit of the PIN. After pressing the enter key ◀, the second digit will flash. Use the programing keys ▲ and ▼ to enter the second digit of the PIN, etc. After the last digit of the PIN, you can use the programing keys ▲ and ▼ to set the target volume. Confirm the selected target volume using the enter key ◀. The parameter P-addit. is set next.

```
▶VT min.....200ml
Exit Menu
```

P-addit. (only in modes S, T and ST in TREND II BILEVEL ST30)

To reach the target volume, an additional pressure (P-addit.) must be set. The pressure is then increased until the target volume is reached, but no further than the additional pressure set. Once the target volume has been reached, the additional pressure is decreased again. The additional pressure is limited to 2 hPa per breath.

If the I-Pressure is set to ≥ 28 hPa then the target volume is automatically set to "OFF". "P-addit." can then no longer be selected.

First set the VT min. Then set the desired value using programing keys ▲ and ▼. To confirm, press the enter key ◀.

```
▶P-addit...3.0hPa
Exit Menu
```

TI min (only in modes S and ST in TREND II BILEVEL ST20 and ST30)

TI min is the minimum inspiration time. By setting of TI min a minimum duration of the inspiration is guaranteed. Breathing out before expire the set TI min is not possible.

To set the minimum inspiration time, first go to the Clinic-menu. Then select **TI min** using the programing keys ▲ and ▼ and press the enter key ◀. You will now be requested to enter the PIN. The first digit will flash. Use the programing keys ▲ and ▼ to enter the first digit of the PIN. After pressing the enter key ◀, the second digit will flash. Use the programing keys ▲ and ▼ to enter the second digit of the PIN, etc. After the last digit of the PIN, you can use the programing keys ▲ and ▼ to enter the minimum inspiration time. Confirm the selected value using the enter key ◀. The TI max parameter is set next.

```
▶TI min.....1.2s
Exit Menu
```

```
▶TI max.....3.0s
Exit Menu
```

TI max (only in modes S and ST in TREND II BILEVEL ST20 and ST30)

TI max is the maximum inspiration time. By setting of TI max a maximum duration of the inspiration is defined. After expiration of TI max the inspiration will be interrupted and the expiration begins.

First set TI min. Then set the desired value using the programing keys ▲ and ▼. To confirm, press the enter key ◀.

```
▶* Low MV.....3.0l
Exit Menu
```

Low MV (only in modes S, T and ST in TREND II BILEVEL ST30)

Here you can set whether the device gives an alarm if the respiratory minute volume falls below the set value.

To set the alarm parameter “Low MV”, first go to the Clinic-menu. Then set the desired value using the programing keys ▲ and ▼. To confirm, press the enter key ◀. If the device is in S mode and the Backup frequency is set to “OFF”, the apnea parameter is set next.

```
▶* Apnea.....2s
Exit Menu
```

Apnea (only in mode S in TREND II BILEVEL ST30, if backup freq. “OFF”)

Here you can set whether the device gives an alarm if an apnea occurs which lasts longer than the set value.

First, set the alarm parameter “Low MV”. Then set the desired value using the programing keys ▲ and ▼. To confirm, press the enter key ◀.

PARAMETER SETTINGS

	Settings range	Dependency	Settings steps	Accuracy
Press.	4 – 20 (30 ¹) hPa	---	0.5 hPa	0.5 hPa
FLEX-Level	1 – 3	---	1 Level	---
P-Start	4 – 20 hPa	\geq P-Min, \leq P-Max	0.5 hPa	0.5 hPa
P-Min	4 – 20 hPa	\leq P-Start	0.5 hPa	0.5 hPa
P-Max	4 – 20 hPa	\geq P-Start	0.5 hPa	0.5 hPa
Mask	Nose, Nose/Mouth	---	---	---
I-Pres.	E-Pres. – 20 (30 ¹) hPa	\leq E-Pres. – 20 (30 ¹) hPa	0.5 hPa	0.5 hPa
E-Pres.	4 hPa – I-Pres.	\geq 4 hPa, \leq I-Pres.	0.5 hPa	0.5 hPa
I-Slope	0.0 – 1,0 s	I-Pres. - E-Pres., \leq I:E	0.1 s	---
E-Slope	0.0 – 1.0 s	I-Pres. - E-Pres.	0.1 s	---
I-Trigger	1 – 3	---	1 Level	---
E-Trigger	1 – 3	---	1 Level	---
Frequency	5 – 30 bpm	---	1 bpm	1 bpm
I:E	0.4 – 8.0 s	I-Slope, E-Slope, Frequency	0.1 s	0.1 s
Delay Time	0 – 20 s	---	1 s	1 s
Backup freq	ON; OFF	---	---	---
VT min ¹	OFF; 200 – 1500 ml	I-Pres. < 28 hPa	100 ml	100 ml
P-addit. ¹	1 – 10 hPa	P-addit. + I-Pres. \leq 30 hPa, I-Pres. < 28 hPa, VT min \geq 200 ml	0.5 hPa	0.5 hPa
TI min	OFF; 0.1 s – (TI max - 0.1 s)	TI max - 0.1 s	0,1 s	0.1 s
TI max	OFF; 0.1 s – 4.0 s	TI min + 0.1 s	0.1 s	0.1 s
Ramp	0 – 60 min	---	1 min	1 min
P-Ramp	4 hPa – E-Pres. APAP: 4 hPa – P-Start CPAP/FLEX: 4 hPa – Press.	---	0.5 hPa	0.5 hPa
Mask Test	0 – 90 s	---	5 s	1 s

¹ TREND II BILEVEL ST30


	Settings range	Dependency	Settings steps	Accuracy
Auto	OFF, Start/Stop, Start	---	---	---
♣ Low MV ¹	OFF, 2 – 10 l/min	---	0.5 l/min	0.5 l/min
♣ Apnea ¹	OFF; 5 – 30 s	Backup freq = OFF	1 s	1 s
Display VT	ON, OFF	---	---	---
Brightness	0 – 100 %	---	10 %	---
Language	DEU, ENG, ELL, SPA, FRA, ITA, NLD, TUR, PLK, POR	---	---	---
P-Unit	hPa, mbar, cmH ₂ O	---	---	---

¹ TREND II BILEVEL ST30

ALARM FUNCTIONS OF THE DEVICE

POWER FAILURE

If a power failure at night would go unnoticed as a result the patient would breathe used air from the therapy tube all night. To avoid this the device is equipped with an alarm mechanism that warns you if power or a fuse fail at night. You will be woken up with an acoustic signal so you can take the mask off your face and breathe fresh air. The alarm is turned off by pushing the on/off key. After power supply is re-established the motor starts automatically and the display will show the following message:



Power Failure
Restart Turbine

Checking the power failure function

IMPORTANT

To check the power failure alarm, the device must have been connected to the power cord and turned on for at least half an hour.

Use the device with running blower and then pull the power plug out of the wall socket. Now the acoustic warning signal should sound. Check once a month if the signal sounds long enough (reference value: at least 3 minutes). After pushing the on/off key or turning on the power supply again the signal will stop.

MASK ALARM

This alarm function is activated only if the automatic function is turned to “OFF” or “Start”. If the mask has slipped off your face or if the tube is pulled off the device, the device is not turned off automatically, but emits a visual and audible alarm. The alarm is silenced after pressing the on/off key or eliminating the leak.



Mask Alarm.....10s
Check Mask!

Check Mask Alarm

Use the device with running turbine in the manual mode (automatic function “OFF” or “Start”). The mask is open but not applied to your face. After a short time, an audible signal is emitted.



Press too high
25.May.11 21:33

PRESSURE ALARM

If during the therapy by a hardware error or other circumstances (e.g. due to a bent measuring tube) the pressure becomes too high the device will generate an alarm sound and turn off the blower. The alarm is turned off by pushing the on/off key and the therapy can be continued after solving the error.



Wake Alarm 07:00
5.0hPa

WAKE ALARM

The device is provided with a comfortable integrated alarm clock. Using the programming keys, you can activate and deactivate this alarm clock. The alarm clock emits an alarm at the programmed wake time. Press the on/off key once to stop the alarm for the next 5 minutes (can be set with the help of the PC Software TRENDset) and twice to deactivate it completely.

The alarm clock function may have been deactivated through the PC Software TRENDset. Please contact your specialized dealer or service technician for activating these functions.



Low MV
8.5hPa

LOW RESPIRATORY MINUTE VOLUME (ONLY ST30)

If the respiratory minute volume falls below the set alarm parameter “Low MV”, the device will give an alarm. The alarm is turned off by pressing the on/off key.



Apnea
8.5hPa

APNEA ALARM (ONLY ST30 IN MODE S)

If an apnea is detected which is longer in duration than the set alarm parameter “Apnea”, the device gives an alarm. The alarm is turned off by pressing the on/off key.

USING AN SD CARD

GENERAL INFORMATION

The use of an SD card during therapy enables the most important therapy data to be saved. The physician can read and evaluate this data via the TRENDset PC software. Furthermore, the physician can set new therapy parameters using TRENDset and send them to the patient. As soon as the SD card is inserted, the device adopts the therapy parameters.

IMPORTANT

SD cards with the specification $1.x \leq 2\text{ GB}$ can be used. The SD card must be formatted with TRENDset before use.

INSERTING THE SD CARD

1. If connected, disconnect the humidifier from the device.
2. Ensure that the device is in standby mode.
3. Carefully lift the device.
4. Insert the SD card into the SD card slot (see picture).
5. The device initializes and the message “Do not remove the SD card” is shown on the display. The time required for initialization depends on the device settings.



NOTICE

On no account remove the SD card during initialization. This could lead to loss of data or damage to the SD card.

6. After initialization, the message disappears. If the SD card contains new therapy parameters, these parameters are adopted by the device and the message “New Parameters were accepted” is displayed.
7. By pressing the enter key ◀ or after 10 s the message will be disappear.



IMPORTANT

After the takeover of the parameters enter the Info-Menu and compare the displayed parameters with the values of your prescription.



Do not remove
the SD card

New Parameters
were accepted



REMOVING THE SD CARD

1. If connected, disconnect the humidifier from the device.
2. Ensure that the device is in standby mode.
3. Carefully lift the device.
4. Carefully push the SD card into the card slot and remove the SD card.

NOTICE

Only remove the SD card in standby mode (see page 24). Otherwise data may be lost or the SD card damaged.

WHAT IS SAVED ON THE SD CARD

The last 100 events are saved on the SD card. Events are changes of parameters such as therapy settings, resetting counters (e.g., filter counters) etc. The parameters are saved with the date and time they were changed, as well as the old and new value. All alarms that occurred are also saved with date and time.

The SD card can only be read using the TRENDset PC software.

The following therapy-related events are also saved on the SD card:

Device type Parameter	CPAP	APAP	BILEVEL	BILEVEL ST20	BILEVEL ST30
Pressure	•	•	•	•	•
Obstruktive Sleep Apnea	-	•	-	-	-
Mixed Sleep Apnea	-	•	-	-	-
Hypopnea	-	•	-	-	-
Apnea	•	-	•	•	•
Airway Constriction	-	•	-	-	-
Snoring	-	•	-	-	-
General Artifact	-	•	-	-	-
Adjustment	•	•	•	•	•

Parameter \ Device type	CPAP	APAP	BILEVEL	BILEVEL ST20	BILEVEL ST30
Mouth Exhalation	-	•	-	-	-
Tidal Volume	•	-	•	•	•
Apnea alarm	-	-	-	-	•
Low MV	-	-	-	-	•

Press ON/OFF!
SD-Error

ERROR MESSAGES

Two different types of error are distinguished: errors with data loss and errors without data loss.

The display “Exx” denotes errors in which data may have been lost. The display “Wxx” denotes errors in which no data has been lost.

To hide an error message, press the enter key ◀.

NOTICE

If error messages occur frequently, we recommend that the device is examined by a service technician.

IMPORTANT

If you carry out the therapy without recording the therapy data, we recommend not using an SD card in the device, in order to avoid unnecessary error messages.

The following table shows all error codes that can occur in connection with the use of the SD card.


Error code	Meaning	Remedy
E00	Writing data has failed. Possible causes are power failure, removal of card during writing or a faulty SD card.	The SD card is probably faulty. Remove the SD card and then insert it into the device again. If the error continues to occur, send the card back to the physician.
E01	The SD card is faulty.	Obtain a new SD card from your physician or specialist dealer.
E02	File system or folder structure is faulty or cannot be read.	The SD card must be formatted by the physician using TREND-set.
E03	New files cannot be created or the SD card is full.	The SD card must be formatted by the physician using TREND-set.

Error code	Meaning	Remedy
E04	The device cannot adopt the therapy data.	The SD card must be returned to the physician to check the therapy.
E05	The SD card was removed.	Insert the SD card again.
	The SD card has fallen out of the card slot.	If this occurs frequently, the device should be examined by a service technician.
E06	The SD card or the device is faulty.	Change SD card. If this occurs frequently, the device should be examined by a service technician.
W00	The device is attempting to repair the SD card. But you have inserted a new SD card into the device.	Confirm with key ◀. The old SD card is defective. Continue to use the new SD card.
W05/06	see error code E05/E06	see error code E05/E06

IMPORTANT DISPLAY MESSAGES

The sections below describe the most important display messages. All further messages will be clear from their context.

You have 45 sec left for seating the mask properly. Once this time has elapsed, the device starts with the soft start function or the therapy.



||||| 8.0hPa
Mask Test.....45s

There will be 10 min and 45 sec left until the full therapy pressure is reached.



||||| 5.0hPa
Ramp.....10:45

You have activated the display of the respiratory volume and the wake up alarm.



||||| 8.5hPa
1.14l *05:30

The mask has slipped off your face or you failed to put the mask on your face while the automatic mode was turned off and the blower was running.



Mask Alarm.....10s
Check Mask!

You pressed the programing key ▲ during operation. There will be a message that you will be awakened at 07.00 o'clock.



Press ON/OFF!
Alarm ON 07:00

There will be 1 min and 18 sec left until the alarm clock will awake you again.



||||| 5.0hPa
Slumber.....1:18

The wake alarm has been initiated. Press the on/off key once to initiate the slumber phase. Press the key once again to turn off the alarm clock completely.



||||| 5.0hPa
Wake Alarm 07:00

You have pressed the on/off key and stopped the slumber phase.



||||| 5.0hPa
Wake Alarm OFF

You pressed the programing key ▼ in the normal operating mode. There will be a message that the alarm clock has been turned off.



Press ON/OFF!
Alarm Clock OFF

The therapy pressure is too high and the turbine switches off.



Press too high
25.May.11 21:33

The automatic function is turned on. You will be requested to put on the mask and breathe.

After the line voltage has been restored, the turbine restarts automatically.

There has been an error during the system start after switching on the voltage supply. **Contact your dealer.**

You have taken down the mask or the mask has slipped off your face while the automatic function was activated. The turbine will turn off after 4 seconds.

The PIN you have entered is wrong. Please enter the correct PIN.

There has been an error during the system start after switching on the voltage supply. **Contact your dealer.**

The Clinic-Menu has been activated.

The Patient-Menu has been activated.

Change filter cassette (see page 63).

Put on Mask!
25.May.11 22:15

Power Failure
Restart Turbine

Watchdog defect

Turbine off.....4s
25.May.11 07:01

Invalid PIN!

MemoryF defect

Clinic-Menu
active

Patient-Menu
active

||||| 5.0hPa
Change Filter

CHANGING THE FILTER CASSETTE, CLEANING


⚠ WARNING

Make sure that you follow the cleaning instructions. If you do not do this, it could damage your health because of germs.

CHANGING THE FILTER CASSETTE

⚠ CAUTION

If the filter is polluted or the display shows the message **Change Filter** the filter cassette must be changed.



To change the filter cassette, proceed as follows:

1. To change the filter cassette, pull it out of the device.
2. Separate filter cassette and filter frame cover by pulling them apart.
3. Replace the filter cassette by a new one. Mount the filter frame cover onto the new filter cassette. Make sure the upper part (slit) and the filter cassette (pin) are in the correct position.

To set back the message proceed as described in the chapter “Patient-Menu” on page 34.

NOTICE

Never operate the device without a filter cassette. Only and exclusively use original HOFFRICHTER filter cassettes.

CLEANING THE MASK

For reasons of hygiene the mask should be cleaned daily:

1. Disconnect the mask from the therapy tube.
2. Clean the mask with mild soapy water. Do not use any other agents!
3. Rinse the mask thoroughly with clear water.
4. Let the mask dry completely in the air.

CLEANING OF THE THERAPY TUBE

For reasons of hygiene the therapy tube should be cleaned weekly. Make sure the measuring tube is unobstructed at all times. Avoid ingress of water into the measuring tube during the weekly cleaning. If water drop-lets should have accidentally entered the measuring tube they must be carefully and thoroughly removed.

1. Close the measuring tube tightly with the tube plugs before cleaning.
2. Clean the therapy tube with mild soapy water. Do not use any other agents!
3. Rinse the therapy tube thoroughly with clear water.
4. Let the therapy tube dry completely in the air.
5. After cleaning, remove both of the tube plugs from the measuring tube.

NOTICE

If water should stay in the measuring tube it could adulterate the pressure measuring or delay the control. Both could affect an optimal therapy.

CLEANING THE DEVICE

For reasons of hygiene, the device should be cleaned weekly:

1. Unplug the power plug.
2. Wipe the device with a cloth slightly dampened with soapy water.
3. Rub the device dry with a cloth.

IMPORTANT

Chemical or thermal disinfecting of the device or the device parts by the user is not necessary.

CLEANING THE HEAD GEAR

Clean the head gear as required:

1. Disconnect the head gear from the mask.
2. Clean the head gear as described in the head gear manufacturer's User's manual.

CLEANING THE HUMIDIFIER

⚠ WARNING

When using a humidifier, the cleaning instructions in the user's manual must be observed.

TROUBLESHOOTING

Problem	Possible cause	Remedy
Pain in the nose, the paranasal sinuses or the ears	Inflammation of the paranasal sinuses or the middle ear	Stop the treatment and contact your physician
Feeling, that the pressure is too high	Malaise with prescribed high pressure values	If you suspect an error, please ask your physician for help
	Acclimatisation phase to the pressure not yet completed	Try to relax. Use or vary the soft start function
Dryness and irritation in the nose and throat	Air is too dry	Device probably does not have an air humidifier. Speak to your physician about retrofitting an humidifier
Original symptoms of sleep apnea come back	Physical condition or life circumstances have changed	Inform your physician
	Device error	Have the device checked by customer services
Irritated or dry eyes	Air escapes between the mask and the skin of the face	Check the mask for proper positioning
		Check the positioning of the mask
Cold nose	Room temperature too low	Replace the mask if the material has become chapped
Runny nose, sneezing	Reaction to the air flow	Either increase the humidity in the room or the temperature of the humidifier
	Normal cold	Contact your physician

Problem	Possible cause	Remedy
Reddening of the skin in the mask area, skin swelling	Incorrect mask size	Inform your physician
	Headgear too tight	Loosen the headgear
	Allergic reaction	Inform your physician
Feeling that the air is too hot	Filter dirty	Change the filter cassette
	Air inlet blocked	Change the position of the device. Keep the air inlet clear
	Heater close to the device	Move the device and the heater further apart
No air flow	Device is defective	Inform customer services
	Water in the measuring tube	Remove the water
Very little air flow	Soft start function has been selected	Reduce soft start time
	Air channels are blocked	Check air inlet
Blower is running constantly at maximum speed	Measuring tube not connected or clogged	Check the measuring tube
	Leak in the device	Have the device checked by customer services
	Water in the measuring tube	Remove the water
Therapy pressure not steady	Water in the measuring tube	Remove the water
	Device defective	Notify the service
	Measuring tube bent	Check for bends

MAINTENANCE

If operated correctly, the device is maintenance-free. However, in order to increase the service life, we recommend regular maintenance every 2 years.

CAUTION

Do not try to open the device. Maintenance and repairs may only be performed by personnel authorized by us, because only they have the tools and measuring equipment required.

You yourself can help to increase the service life of the device and ensure that it continues to work safely.

- Follow the cleaning instructions from page 63.
- Check the system regularly:
 - Conduct a visual check for external damage and dirt
 - Check the alarm function „Power failure“ (see page 54)
 - Check the alarm function „Mask alarm“ (see page 54)

PREPARING THE DEVICE WHEN CHANGING PATIENT

The devices are only intended for use by a single patient.

CAUTION

If the device is to be used for another patient, it must first of all be prepared hygienically.

When being given to another patient, the device must be prepared hygienically by the specialist dealer or the manufacturer. If reuse of the mask and the therapy tube is planned, they must also be prepared by the specialist dealer or the manufacturer.

The preparation procedure is described in detail in the corresponding hygiene plan.

DISPOSAL

Proper disposal saves natural resources and prevents harmful substances being released into the environment.

DEVICE



The device must not be disposed of with the household waste. Please contact the relevant customer services department to find out how to dispose of the device properly.

PACKAGING



The packaging is taken back by the distributor but it can alternatively be recycled.

ACCESSORIES

The accessories like the tubing, mask, filter cassettes, etc. should be disposed according to the manufacturer's instructions, or with normal household waste.

ACCESSORIES

Scope of delivery	Article number
TREND II CPAP	0000 2061
TREND II AUTO	0000 2064
TREND II BILEVEL	0000 2062
TREND II BILEVEL ST20	0000 2063
TREND II BILEVEL ST30	0000 2035
Carrying case	0000 2437
Power supply cable (approx. 2.30 m)	3110 0015
Filter cassette	0000 2058
Therapy tube (1.80 m), measuring tube and tube plugs	0000 7116
SD card, formatted with TRENDset	1120 0010
User's manual	5000 0427
Brief instruction	5000 0434

Optional	Article number
Silicone CPAP nasal mask, size S	0000 3440
Silicone CPAP nasal mask, size M	0000 3434
Silicone CPAP nasal mask, size L	0000 3435
Silicone CPAP full face mask, size S	0000 3441
Silicone CPAP full face mask, size M	0000 3436
Silicone CPAP full face mask, size L	0000 3437
ComfortTube System (heated tubing system)	0000 3479
Humidifier AquaTREND uni	0000 2073
Filter cassette with frame cover	0000 2057
24 V DC vehicle cable	0000 9212
12 V to 24 V converter	0000 7133

For ordering of accessories, please contact a HOFFRICHTER service partner.

TECHNICAL DATA

	TREND II					TREND II with AquaTREND uni				
Dimensions	215 x 195 x 105 (W x D x H) mm					215 x 320 x 125 (W x D x H) mm				
Weight	2.3 kg					2.65 kg (without water)				
Power supply	100 V~ ... 240 V~, 50 ... 60 Hz, 24 VDC / 2.1 A									
Mains power	100 ... 240 VAC, 50 ... 60 Hz									
DC power	24 VDC / 2.1 A									
Operating time under battery power	up to 8 years									
Pressure range	4...20 hPa (mbar); BILEVEL ST30: 4...30 hPa (mbar)									
Max. limit pressure in the event of a fault	≤ 30 hPa (CPAP - APAP) ≤ 40 hPa (BILEVEL - BILEVEL ST20 und ST30)									
Power consumption	DC power		Mains power			DC power		Mains power		
Standby	< 5 W		< 5 W			---		---		
Operation at 12 hPa	< 10 W		< 12 W			< 30 W		< 33 W		
Short term pressure variation	4 hPa	10,5 hPa	17 hPa	23,5 hPa	30 hPa	4 hPa	10,5 hPa	17 hPa	23,5 hPa	30 hPa
10 bpm	0.15 hPa					0.16 hPa (at heating level 3)				
15 bpm	0.30 hPa					0.30 hPa (at heating level 3)				
20 bpm	0.46 hPa					0.49 hPa (at heating level 3)				
Long term pressure variation	± 0.3 hPa									
Pressure reading accuracy	0.5 hPa									
Average sound pressure level (operating at 1 m distance)	≤ 26 dB(A) at 10 hPa (equivalent to a sound power level of ≤ 34 dB[A])									
Air flow rate										
4 hPa	> 180 l/min									
7.5 hPa	> 170 l/min									
15 hPa	> 160 l/min									
22.5 hPa	> 150 l/min									
30 hPa	> 140 l/min									
Operating temperature	+ 5 °C ... + 40 °C									

	TREND II	TREND II with AquaTREND uni
Storage temperature	- 20 °C ... + 70 °C	
Therapy air heating	< 3 K	depending on the heating level and the ambient conditions
Relative humidity	15 % ... 95 % for operation and storage	
Operating conditions	ca. - 400 m ... 3500 m (1060 hPa ... 700 hPa)	
Filter	Polyurethane foam on polyester base, 80 ppi, 30 kg/m³	
Therapy tube connection	22 mm, cone (gem. ISO 5356-1)	
Product class according to 93/42/EEC	IIa	
Classification according to EN 60601-1	Protection class II	

Therapy parameter factory settings	Device type				
	CPAP	AUTO	BILEVEL	BILEVEL ST20	BILEVEL ST30
Mode	FLEX	APAP	S	ST	ST
Press.	6,0 hPa	---	---	---	---
FLEX-Level	2	---	---	---	---
P-Start	---	6,0 hPa	---	---	---
P-Min	---	4,0 hPa	---	---	---
P-Max	---	10,0 hPa	---	---	---
Mask	---	Nose	---	---	---
I-Pres.	---	---	10,0 hPa	10,0 hPa	10,0 hPa
E-Pres.	---	---	6,0 hPa	6,0 hPa	6,0 hPa
I-Slope	---	---	0,3 s	0,3 s	0,3 s
E-Slope	---	---	0,3 s	0,3 s	0,3 s
Trigger	---	---	---	---	---
I-Trigger	---	---	2	2	2
E-Trigger	---	---	2	2	2
Frequency	---	---	---	12 bpm	12 bpm
I:E	---	---	---	2,0 s	2,0 s
Delay Time	---	---	---	5 s	5 s
Backup freq	---	---	OFF	---	---

Therapy parameter factory settings	Device type				
	CPAP	AUTO	BILEVEL	BILEVEL ST20	BILEVEL ST30
VT min	---	---	---	---	OFF
TI min	---	---	---	OFF	OFF
TI max	---	---	---	OFF	OFF

Alarm parameter factory settings TREND II BILEVEL ST30

Low MV	OFF
Apnea	OFF

Comfort parameter factory settings for all device types

Ramp	0 min
Mask Test	0 s
Auto	OFF
Display VT	OFF
Brightness	50 %
P-Unit	hPa

CE marking as per EC directive 93/42/EEC.

The manufacturer reserves the right to make technical changes without notice.

MANUFACTURER'S DECLARATION ON ELECTROMAGNETIC COMPATIBILITY

Guidance and manufacturer's declaration – electromagnetic emissions

The TREND II is intended for use in the electromagnetic environment specified below. The user¹ of the TREND II should assure that it is used in such an environment.

Emissions test	Compliance	Electromagnetic environment – guidance
RF emissions acc. to CISPR 16-1-2	Group 1	The TREND II uses RF energy only for its internal function. Therefore, RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF emissions acc. to CISPR 16-1-2	Class B	The TREND II is suitable for use in all establishments including those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.
Harmonic emissions acc. to IEC 61000-3-2	Class A	
Voltage fluctuations / flicker emissions acc. to IEC 61000-3-3	Complies	

Guidance and manufacturer's declaration – electromagnetic immunity

The TREND II is intended for use in the electromagnetic environment specified below. The user¹ of the TREND II should assure that it is used in such an environment.


Immunity test	IEC 60601 Test level	Compliance level	Electromagnetic environment – guidance
Electrostatic discharge (ESD) acc. to IEC 61000-4-2	± 8 kV air	± 8 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30 %.
Electrical fast transient/burst acc. to IEC 61000-4-4	± 1 kV input/output	± 1 kV input/output	Mains power quality should be that of a typical commercial or hospital environment.
Surge acc. to IEC 61000-4-5	± 1 kV voltage differential mode	± 1 kV voltage differential mode	Mains power quality should be that of a typical commercial or hospital environment.

¹ Here user is meant in the sense of “Responsible Organization”

Guidance and manufacturer's declaration – electromagnetic immunity

Immunity test	IEC 60601 Test level	Compliance level	Electromagnetic environment – guidance
Voltage dips, short interruptions and voltage variations on power supply input lines acc. to IEC 61000-4-11	<p>100 % dip in U_T for 0.5 cycle</p> <p>60 % dip in U_T for 5 cycles</p> <p>30 % dip in U_T for 25 cycles</p> <p>100 % dip in U_T for 5 s</p>	<p>100 % dip in U_T for 0.5 cycle</p> <p>60 % dip in U_T for 5 cycles</p> <p>30 % dip in U_T for 25 cycles</p> <p>100 % dip in U_T for 5 s</p>	Mains power quality should be that of a typical commercial or hospital environment. If the user of the TREND II requires continued operation during power mains interruption, it is recommended that the TREND II is powered from an uninterrupted power supply (UPS) or a battery.
Power frequency (50/60 Hz) magnetic field acc. to IEC 61000-4-8	3 A/m	3 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.
Conducted RF acc. to IEC 61000-4-6	<p>$V_1 = 3$ V</p> <p>150 kHz – 80 MHz</p>	3 V	<p>Portable and mobile communications equipment should be used no closer to any part of the TREND II, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.</p> <p>Recommended safety distance:</p> $d = 1.17 \sqrt{P}$

Guidance and manufacturer's declaration – electromagnetic immunity

Radiated RF acc. to IEC 61000-4-3	$E_1 = 3 \text{ V/m}$ 80 MHz – 2.5 GHz	3 V/m	$d = 1.17 \sqrt{P}$ for 800 MHz to 800 MHz $d = 2.33 \sqrt{P}$ for 800 MHz to 2.5 GHz with P as the rated maximum output power of the transmitter in watts (W), according to the transmitter's manufacturer, and d as the recommended safety distance in meters (m). Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey ^a , should be less than the compliance level in each frequency range ^b . Interference may occur in the vicinity of equipment marked with the following symbol. 
-----------------------------------	---	-------	---

- Note 1 U_T is the mains alternating current before application of the test level.
- Note 2 At 80 MHz and 800 MHz the higher frequency range is essential.
- Note 3 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.
- ^a The field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength outside the location in which the TREND II is used exceeds the compliance level, the TREND II should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as re-orienting or relocating the TREND II.
- ^b Over the frequency range from 150 kHz to 80 MHz the field strength should be lower than 10 V/m.

Recommended separation distances between portable and mobile RF communication equipment and the TREND II

The TREND II is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The user¹ of the TREND II can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the TREND II as recommended below, according to the maximum output power of the communications equipment.

Rated maximum output power of transmitter (W)	Separation distance according to frequency of transmitter (m)		
	150 kHz – 80 MHz $d = 1.17 \sqrt{P}$	80 MHz – 800 MHz $d = 1.17 \sqrt{P}$	800 MHz – 2.5 GHz $d = 2.33 \sqrt{P}$
0.01	0.12	0.12	0.23
0.1	0.37	0.37	0.74
1	1.17	1.17	2.33
10	3.7	3.7	7.37
100	11.7	11.7	23.3

For transmitters rated at a maximum output power not listed above, the recommended separation distance d in meters (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

Note 1 At 80 MHz and 800 MHz the higher frequency range is essential.

Note 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

¹ Here user is meant in the sense of “Responsible Organization”

DISCLAIMER

HOFFRICHTER GmbH is not liable for consequences in terms of safety, reliability and performance of the product where:

- interventions, modifications, extensions, calibration, repairs and maintenance are carried out by persons not authorized by us,
- other manufacturers' accessories and spare parts are used that have not been approved by us for use on the product,
- the product is used other than as described in the user's manual or
- the hygiene and cleaning instructions described in the user's manual have not been complied with.

Statutory guarantee rights remain unaffected by this.

.....

.....

.....

.....

.....

.....

.....

.....

.....

.....

.....

.....

.....

.....

.....

.....

.....

.....

.....

.....

HOFFRICHTER GmbH
Mettenheimer Strasse 12/14
19061 Schwerin
Germany
Phone: +49 385 39925-0
Fax: +49 385 39925-25
E-mail: info@hoffrichter.de
www.hoffrichter.de